



BID CLARIFICATION MATRIX NO. 3

**PROCUREMENT OF SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, TRAINING AND AFTER-SALES
SERVICES FOR LABORATORY EQUIPMENT**

TENDER NUMBER: PRQ20241185

JULY 3, 2025

FUNDED BY

French Development Agency



CLARIFICATION MATRIX NO. 3

Procurement of Supply, Delivery, Installation, Commissioning, Training and After-sales Services for Laboratory Equipment

PRQ20241185

This Bid Clarification Matrix No. 3 is an integral part of the solicitation document **and, as such, should be carefully reviewed** by service providers when preparing and submitting bids.

All communications must be addressed to the procuring entity, TradeMark East Africa, through procurement@trademarkafrica.com as outlined in the tender document.

QUESTIONS AND ANSWERS

S/No.	Questions	Clarification Response	Status
1.	<p>Kindly note that there is a slight discrepancy between the list of goods (pages 77–80) and the corresponding list of specifications (pages 105–124) for Lot 4. Specifically, the numbering of several items (highlighted in red) appears to be misaligned between the two lists.</p> <p>Additionally, we have observed that certain items listed in the specifications—such as the Emergency Eyewash and Shower Stations (Item No. 16) and the Filter (Item No. 26)—do not appear in the list of goods.</p> <p>We would appreciate your clarification on this matter to ensure proper alignment between the goods and their specifications.</p>	<p>Follow the numbering in the equipment list provided under Section 1. List of Goods and Delivery Schedule</p> <p>The products to be supplied are only those listed under the Section 1. List of Goods and Delivery Schedule for the specified lots. Bidders not to respond to Emergency Eyewash and Shower Stations (Item No. 16) and the Filter (Item No. 26) listed under specifications but do not appear in the list specified.</p>	Closed
2.	<p>LOT 2. MASS METROLOGY EQUIPMENT</p> <p>Considering the technical specifications sheet provided in the Solicitation document and Clarification Matrix 1.</p> <p>Accessories for OHAUS Adventurer, Explorer and Pioneer balances: Kindly information regarding the Pan size and the</p>	<p>Please consider only the accessories provided in clarification 1 (No. 2 Lot 2, i)</p>	Closed

S/No.	Questions	Clarification Response	Status
	capacity of the Balances.		
3.	<p>Lot 3: line item 5: Digital thermometer with LCD display Multi-channel Thermometer Temperature - Pt100 Measuring range -200 to +800 °C Accuracy ± 0.2 % of mv (Remaining Range) ± 0.2 °C (-100 to +199.9 °C) Resolution 0.05 °C Up to 3 connectable sensors</p> <p>Must be supplied with: Temperature - TC Type K air probes: Measuring range: -60 to +400 °C Accuracy: Class 2 Reaction time: 25 s Dimensions: length: 255 mm Quantity: 5 robust air probes The total quantities of Digital thermometer with LCD display requested are five pieces, Under specifications as shown above the probes quantity is indicated as 5 robust air probes.</p> <p>Kindly clarify whether the total number of air problems needed is five OR each Digital thermometer with LCD display requested, should come with five probes giving a total of 25 robust air probes.</p>	5 thermometers each with 3 sensors or temperature probes covering the specified temperature ranges.	Closed
4.	Is calibration with recognized local body for balances	Calibration by a laboratory with valid accreditation to ISO/IEC 17025 standard	

S/No.	Questions	Clarification Response	Status
	required? If local calibration is required, is this sufficient as <i>valid calibration certificates from an accredited facility conforming to ISO/IEC 17025 Standard?</i>	will be sufficient.	Closed
5.	Is user training required for all items? If training is required, is it in English or French?	Training to be undertaken primarily in French.	Closed
6.	Lot 4, Item 23: Power Cord – You have indicated “Each power cord must be compatible with the equipment offered. Suppliers must identify the corresponding device and provide appropriate cords (e.g., EU/IEC standards) for each” All the offered equipment in this lot are supplied with their power cords. Confirm bidders should not offer additional cords under this item.	The power cord supplied with the equipment by the manufacturer will be considered sufficient. A Type C adapter, commonly known as a Europlug, is preferred.	Closed